



Government of West Bengal  
Office of the Principal  
Nil Ratan Sircar Medical College  
Kolkata – 700014; Phone: 033-2286-0033  
Website: [www.nrsmc.edu.in](http://www.nrsmc.edu.in); e-mail: [principalnrsmc@gmail.com](mailto:principalnrsmc@gmail.com)

Memo No. NMC/STORE/ 4274

Date: 31/08/2019

**Re-E-TENDER NOTICE (NIT No: 30/ 2019-20) FOR PURCHASE AND INSTALLATION OF  
EQUIPMENTS FOR DIFFERENT DEPARTMENT OF  
NIL RATAN SIRCAR MEDICAL COLLEGE, KOLKATA - 700014.  
(Through Pre-qualification)  
(Submission of Bid through NIC e- tender portal)**

The Principal, Nil Ratan Sircar Medical College, Kolkata – 700014, is going for e-Tender on behalf of the Health & Family Welfare Department, Government of West Bengal for obtaining the rate of equipment as per list from bonafide Manufacturer/Direct Importer/Authorized Distributors. Necessary earnest money to be submitted for participation in the Tender is appended in the table below. Exemption of EMD is applicable to agencies having valid, item specific certificate issued by appropriate authority of Govt. of West Bengal.

1)

Serial No	Name of equipment	Department	Quantity	Earnest Money
1	HPLC	MRU	01 pc	Rs. 20000.00

2)

Serial No	Name of equipment	Department	Quantity	Earnest Money
1	Elisa reader and Washer	MRU	01 pc	Rs. 20000.00

3)

Serial No	Name of equipment	Department	Quantity	Earnest Money
1	Fluorescence Microscope	MRU	01 pc	Rs. 20000.00

4)

Serial No	Name of equipment	Department	Quantity	Earnest Money
1	Gastro Video scope & Colono video scope with Video processor & light Source.	Gastroenterology unit under General Medicine	01 pc	Rs. 10000.00

5)

Serial No	Name of equipment	Department	Quantity	Earnest Money
1	Duodeno Video scope with Video processor & light Source	Gastroenterology unit under General Medicine	01 pc	Rs. 10000.00



6)

Serial No	Name of equipment	Department	Quantity	Earnest Money
1	Anesthesia Work station	Anesthesiology	10 pc	Rs. 30000.00

All bidders must complete their technical demonstration of their quoted items within ten(10) days from the date of opening of technical bid to the concern Department.

#### 1. General Instructions:

In the event of e-filing, intending bidder may download the tender documents free of cost from the website: <http://wbtenders.gov.in> directly with the help of Digital Signature Certificate or from the Health & Family Welfare Department's website [www.wbhealth.gov.in](http://www.wbhealth.gov.in).

#### 2. Submission of Bids:

Both Technical Bid and Financial Bid are to be submitted concurrently duly digitally signed, in the website <http://wbtenders.gov.in>. All papers must be submitted in English language.

#### 3. Time Schedules:

The time schedule for obtaining the bid documents, pre bid meetings, registration with the tendering authorities, the submission of bids and other documents etc. will be as per the list provided in clause no 23 as given below.

#### 4. Eligibility for Quoting:

Only Manufacturers/direct Importers/Authorized Distributors are eligible for quoting. The price is to be quoted in Indian Rupees including cost of insurance, packing, forwarding, freight charges, clearing charges and installation of the new instrument and exclusive of GST and other taxes which are to be quoted separately in the BOQ. Agency commission is not permissible.

#### 5. Submission of Tender:

The tender is to be submitted in Two Bid System (Technical Bid or Bid A & Financial Bid or Bid B).

(A) Technical Bid or Bid A:

#### "BID A": PART I

File containing Technical Bid only - single file with multiple scanned pages containing Statutory and My documents separately.

(a) STATUTORY Cover containing the following documents:

(SINGLE FILE MULTIPLE PAGE SCANNED)

Essential Requirements of the Tendering Firm for participation" shall contain all papers related to the essential requirements of the Renderer for participation in the tender viz.

Sl. No.		Page no
1	Duly filled in CHECK LIST in the prescribed format. Page number of documents must be mentioned in the check list before uploading the same.	
2	Application in the prescribed format given in Annexure I	
3	Authorization letter of signatory from Company in Annexure II	
4	Copy of agreement between the manufacturer and the Distributor in Annexure III	
5	No conviction certificate in the form of an Affidavit as per Annexure-IV	
6	A Proof of submission of applicable END	



**(b) NON STATUTORY/MY DOCUMENTS** Cover Containing the following documents:  
Essential Requirements of the Tendering Firm for participation” shall contain all papers related to the essential requirements of the Renderer for participation in the tender viz.

Sl. No.	Category	Sub Category	Sub Category Description	Page No
A	Certificates	A1. Certificates	PAN Card of the authorized signatory	
			Professional Tax Registration certificate	
			PUC in respect of the manufacturing unit	
			GST Registration certificate	
B.	COMPANY DETAILS	B1. COMPANY DETAILS 1	Certificate of country of origin	
			Registration with Registrar of Companies	
			Trade License in the name of participating firm	
C.	CREDENTIAL	C1: CREDENTIAL 1	Manufacturing License with validity	
			Import License with validity (for importers)	
		C2 : CREDENTIAL 2	Manufacturer’s guarantee(in case of Distributors and Direct Importer, to be obtained from manufacturer)	
			A declaration in respect of existence of a fully equipped service centre in or around Kolkata in respect of the relevant item. List of purchaser in Govt. Institutions of repute, for last two years	
D.	FINANCIAL INFO	PAYMENT CERTIFICATE 1	Income Tax Returns submitted for the last three years	
		PAYMENT CERTIFICATE 2	VAT/CST Returns (of the last quarter) for the last three years.	
		Audited P/L & BALANCE SHEET		
		P/L & Balance sheet of the last three years		

**BID A”: PART II**

Technical specifications of Instrument (in PDF/Excel format. Must mention Make, Model, exact features of the offered item against each point given in the specification chart). **Improper filling &/or suppression of facts/false statement will lead to disqualification and penalty in the form of permanent black listing.**

**(B) Financial Bid B**

**COVER “B”:** Separate file containing financial document only.

(a) Financial Bid: BOQ in INR (in excel sheet)

The folder marked as “Financial Bid” shall contain ALL INCLUSIVE PRICE in INR only and should include base price along with the cost of allied works and services which are to be undertaken (like



installation, commissioning of equipment or PC etc.), cost of insurance, packaging, forwarding, delivery charges, clearing charges of the new instrument to be purchased. **GST and any other applicable tax. should be quoted separately. Five years CAMC rate should also be quoted separately** in the template for Bill of Quantities (BOQ). However, the L1 will be determined on base price plus admissible taxes only. GST quoted will be directly remitted to Government account as per existing norms of Finance Department.

The tenderers are required to submit the hard copies of Cover A (Transaction document of EMD , General Documents and Technical Specifications) as per schedule described in the time line vide clause 23, preferably by **speed post/ courier**, to the office of the Principal, NRS Medical College, Kolkata. Submission of hard copy of Cover A to the office of the Principal, Nil Ratan Sircar Medical College, Kolkata, is for reference only while examining the technical bid on line and will not be treated as substitute for on line submission. **Submission of hard copy of Bid B is totally prohibited.**

#### **6. Evaluation of Tender:**

During tender evaluation process, "Bid A" will be opened first. Those tenderers who would qualify the Statutory & Non-Statutory requirements on the basis of technical and commercial documents will be identified and "Bid B" i.e. financial bid of only those qualified bidders will be opened. Verification of hard copies of the audited balance sheet and Profit & Loss Accounts of the Tenderers for the last three years will be made if deemed necessary before opening the Financial Bid (Bid B) of the technically qualified bidders. If found suitable on the basis of above pre-qualification, the tenderer quoting the lowest rate, will be considered as successful.

#### **7. Earnest Money:**

The earnest money of the tenderer is liable to be forfeited if the tenderer withdraws his tender as a whole or for any particular item or items at any stage after opening of the tender or fails/refuses to enter into written agreement for any or all of the items of his accepted bid within the time specified when requested to do so or fails to furnish Performance Bank Guarantee within the stipulated time. **EMD should be deposited online in the Pooling Account of Government of West Bengal maintained at ICICI Bank, RN Mukherjee Road Branch in the manner as prescribed in Government Order no. 3975-F(Y) dated 28/07/2016. (Order copy is available at the site-[www.wbfin.nic.in](http://www.wbfin.nic.in))** Hard copy of Bid A should reach the office on or before the stipulated date as mentioned in clause 23.

#### **8. Rate:**

The price is to be quoted in INR only including cost of insurance, entry tax, packaging, forwarding, delivery charges, clearing charges and installation of the new instrument. The basic rate should be furnished inclusive of Transportation Cost, Insurance, Testing Charges, Incidental Charges etc. but excluding GST and any other applicable taxes etc. which should be quoted separately in the template for Bill of Quantities (BOQ).

#### **9. Order & Supply:**

The tendering authority has the liberty to ask for **performance demonstration** of any item during any stage of the tender process, before placing order for supply. In case of dissatisfaction, the authority will have the right to disapprove the offer of L1 bidder and go for the next bidder. Orders for supply of the approved products will be placed with the successful tenderers after execution of agreements and such supply shall have to be made in such installments as may be fixed or spread over the period to be



specified in the supply orders to be made in pursuance of the agreements. The equipment is to be transported in such packaging so as to avoid damage to primary package of manufacturer. The successful tenderer will have to supply within the specified time schedule that had been assured at the time of selection as supplier.

**The tendering authority will be at liberty to enhance the quantity of the item tendered up to 40% of the original required quantity. Bidder will have to consider the same while bidding as supply of the enhanced quantity will be mandatory as the original one.**

**10. Withdrawal /Cancellation & Purchase Policy of Tendering Authority:**

The tendering authority reserves the right to withdraw any item from the tender at any stage. The selection of such item, if already made in favour of any tenderer, shall be treated as cancelled. The tendering authority reserves the right to reject or accept any tender or part thereof at any stage or to split any tender without assigning any reason. Withdrawal of tender or any revision after submission of tender by the tenderer will not be allowed.

Purchase will, however be made following the existing purchase policy of the Govt. of West Bengal and its amendment(s) made from time to time. The purchase policy of the State Government as provided in the West Bengal Financial Rules, the policy of price preference in particular incorporated under Notification No. 10500-F dated 19.11.04, 5400-F(Y) Dated 25/06/2012 and subsequent relevant orders should be observed in considering the tenders.

**11. Important Instruction with regard to Submission of Tender:**

The rate should be quoted in Indian Rupees for a particular brand or model (model no. of the offered item only must be mentioned in the appropriate column of the 'Tender Form'. Alternative offer will not be accepted). The tender shall be cancelled for the quotation of item without its brand name, model, model no. etc. Detail information regarding the manufacturer and the item quoted in the 'Tender Form' should be furnished.

**12. Spare Parts:**

The Bidder will undertake that supplies of necessary maintenance equipment and spare parts will be made available for all items/equipment and the complete system for at least five years on a continuing basis. However, this does not relieve the supplier of any warranty obligations under the contract.

**13. Guarantee / Warranty Period:**

i) The machine/Instrument should be covered by free on-site comprehensive Warranty for one year after successful installation. The tenderers must quote for 5 (Five) years' comprehensive on-site AMC of entire system (including all spares & labour) after the date of completion of free warranty period. The CAMC charges for 5 years to be quoted in the BOQ, will however be determined on extant Govt. rules in this regard.

ii.) All faults appearing and their rectification shall be periodically advised to the institute, the period being not more than two weeks.

iii.) Any lacuna or lacunae noticed in the functioning of the installation as a result of any design or other feature shall be rectified by the supplier free of cost within 72 hours of call log.

**14. After Sales Services and Maintenance Contract:**

After sales service must be provided by the supplier, during and after warranty period of the equipment. Selected vendor will remain liable to attract penalty for violation of machine downtime as per agreement. The selected vendor or his authorized servicemen has to attend the machine within 72 hours of Call Log. It is desirable to have a fully equipped service centre in or around Kolkata.

**15. Delivery, Installation and Commissioning**

i. Delivery of the goods at the purchaser's premises shall be completed by the Supplier in accordance with the terms specified by the purchaser.



- ii. The installation, testing and commissioning of the proposed system shall be completed in accordance with the order.
- iii. All items are to be delivered within six weeks from the date of placement of order.
- iv. In case of distributor, the firm should be direct distributor of manufacturing unit. **Sub distributor authorization by distributor will not be accepted at all.**

**16. Penalty Clauses:**

**PENALTY FOR FORMATION OF CARTEL OR FURNISHING OF FRAUDULENT/ MISLEADING DOCUMENTS:** If during the tender process or at any stage during the validity of the tender period, it is found that a tenderer(s) has formed a cartel in what so ever form or name to fix up the rates or suppliers to the detriment of the fairness of the tender process, penal measures shall be initiated. Similar penal measures shall also be initiated against those tenderers who have submitted false/ misleading/ fraudulent documents or made incorrect declarations. The penal measure will be forfeiture of Earnest Money, forfeiture of Performance Bank Guarantee if enlisted as a supplier in addition to any other legal action against the firm.

**17. Appeal:**

Appeal against the decision of the Principal, Nil Ratan Sircar Medical College, Kolkata and then to impose such a penalty will lie with Tender Selection Committee. Review against the decision of the Tender Selection Committee (TSC) will lie with Health & F.W. Dept. of the Govt. of W.B. The Special Secretary/Secretary/Principal Secretary will be the appellate authority within the Department of Health & Family Welfare, Government of West Bengal. Before imposing any penalty as per clauses 20, 21 and 22 the concerned supplier may appeal to the authority citing the proper reasons for non- imposing the penalty as stated.

**18. Agreement:**

On a tender being accepted, intimation of acceptance will be forwarded to the tenderer by Principal, Nil Ratan Sircar Medical College, Kolkata. After communication of the same, the tenderer will have to execute agreement in the prescribed form with the Principal, Nil Ratan Sircar Medical College, Kolkata. In case any direct purchasing unit wishes to go for a separate agreement, the head of the purchasing unit must get written permission to that effect from his controlling authority. This present document and the tender forms filled in by the tenderer or copies thereof in so far as they are not inconsistent with these terms & conditions will be incorporated as part of the agreement. Such agreement will be binding on the tenderer.

**19. Validity Period of Agreement:**

The contract period will be for a period 01 (one) years from the date of finalization of tender **which can be extended for a further period of 6 months if Tender Selection Committee feels necessary.**

**20. Performance Bank Guarantee:**

The performance bank guarantee will be mandatory for all suppliers and will not be waived in any case. The successful tenderers shall be required to furnish the 'Performance Bank Guarantee' from any Nationalized/Scheduled Bank acceptable to Govt. of West Bengal for One year @ 10% of quoted base rate per item for which the tenderer has been selected as supplier. The performance bank guarantee will have to be renewed after completion of Free Warranty period, for additional 5 (Five) years under the paid CAMC period before transfer of the CAMC amount to bidder's account. Mentioning of West Bengal Government Pooling Account for Performance Bank Guarantee No. 000605030134 IFS Code ICIC0000006 MICR 70229002 at ICICI Bank, RN Mukherjee Road Branch, Kolkata on the body of Performance Bank Guarantee is made mandatory vide Finance Department GO No. 2691-F(Y) dated 02/05/2017.

**21. Inspection:** Before submitting the tender, the intending tenderers should thoroughly get acquainted themselves with the proposed supply and installation by local inspection of site and take into



consideration the site condition and other criterion for effecting smooth supply. No claim whatsoever will be entertained afterwards.

## 22. Payment Terms:

Payment will be made electronically by Govt. of West Bengal approved system after execution of due supply and successful installation as ordered, subject to :

- i. Submission of Performance Bank Guarantee in terms of Clause 20 and subject to penalty clause in terms of Clauses 16.
- ii. Supply of the materials as per specification as provided in the tender documents and the catalogue. Supply of the materials must be made within the stipulated period as specified in the work orders.
- iii. On being selected, the successful vendors will have to submit one application to the Principal, Nil Ratan Sircar Medical College, Kolkata stating the name of the payee/recipient, Bank account no with IFS code of the payee/recipient for making e-payment. Payment will only be released after satisfactory installation and verification of following items-
  1. Supplier's invoice in original
  2. Challan and installation certificate
  - 3 Any other thing of importance on the discretion of College Authorities

## 23. Dates & Information:

S. No	Items	Dates
1	Date of uploading of N.I.T. Documents (online) from this end.	04.09.19, 9.00 A.M.
2	Documents download start date (Online)	04.09.19, 10.00 A.M.
3	Date of hoisting of documents at Departmental Website ( <a href="http://www.wbhealth.gov.in">www.wbhealth.gov.in</a> )	04.09.19, 9.00 A.M.
4	Date of Pre Bid Meeting and workshop with the intending bidders at the office of the Principal of NRS Medical College	06.09.2019, at 1.00 P.M
5	Bid Submission starting (On line)	07.09.2019 9.00 A.M. onwards
6	Bid Submission closing (On line)	21.09.2019 at 4 P.M.
7	Bid opening technical bid(online) (Bid A )	23.09.2019 at 4.00 P.M.
8	Date of uploading list for Technically Qualified Bidder (online)(Bid A)	24.09.2019 10 A.M. onwards.
9	Date & for opening of Financial Proposal (Bid B) (Online)	To be declared later
10	Date of uploading of list of bidders along with the approved rate	To be declared later.

Principal, Nil Ratan Sircar Medical College, Kolkata, reserves the right to change the above schedule in case of any exigencies after putting up a notice in the departmental website and Notice Board of N R S Medical College, Kolkata.

**24. Validity of Tender:**

- i. The tender will remain valid up to 1 (one) year from the date of finalization of tender **which can be extended for a period of 6 months**. Any objection raised by any bidder in this respect will not be entertained. No informal tenderer will be entertained in the bid further.
- ii. During the scrutiny, if it comes to the notice of tender inviting authority that the credential or any other paper found incorrect/ manufactured/ fabricated, that bidder would not be allowed to participate in the tender and that application will be out rightly rejected without any prejudice.
- iii. Prospective bidders must take training from authorized agencies before submission of their bids.
- iv. The Tender Selection Committee reserves the right to cancel the N.I.T. due to unavoidable circumstances and no claim in this respect will be entertained.

**25. NO- CONVICTION CERTIFICATE:**

No conviction certificate in the form of An Affidavit is to be submitted in the prescribed Pro-forma as Annexure-IV attached herewith from Notary Public/ Judicial Magistrate/Executive Magistrate.



Principal  
NRS Medical College  
Kolkata-14



**CHECK LIST (MUST BE PROPERLY FILLED)**

**NOTE: For Non applicability for some documents please mention "NA" in the check list.  
Must be mention page number in the check list.**

S. No	Item	Pl put tick or NA	Page No.
1	Application submitted in Annexure I		
2	Annexure II (Authorization letter of signatory from the Company)		
3	Annexure III (Copy of agreement between the Manufacturer and the Direct Importer/Distributor)		
4	No conviction certificate in the form of an Affidavit as per Annexure-IV.		
5	Income Tax Returns submitted for the last three Assessment years		
6	GST Registration certificate(not provisional)		
7	PAN Card of the organization and authorized signatory.		
8	Manufacturer's guarantee (in case of Distributors and Direct Importer, to be obtained from manufacturer)		
9	Audited P/L & Balance sheet of last three years		
10	Technical specifications of Instrument (in PDF/Excel format. Must mention Make, Model, exact features of the offered item against each point given in the specification chart).		
11	Trade License in the name of participating firm		
12	PUC in respect of the manufacturing unit		
13	Certificate of country of origin		
14	Registration with Registrar of Companies		
15	Valid Manufacturing License		
16	Import License with validity (for importers)		
17	List of purchaser in Govt. Institutions of repute, for last two years		
18	A declaration in respect of existence of a fully equipped service centre in or around Kolkata in respect of the relevant item.		
19	A Proof of submission of applicable END		



**ANNEXURE-I**

**APPLICATION FORMAT**

(To be furnished in the Company's official letter pad with full address and contact no etc)

To The Principal, Nil Ratan Sircar Medical College, Kolkata-700014.

Sub: Tender No.-----Dt.------(Name of the Item -1)-----2)-----  
-----3)-----4)-----

----) for purchase and installation of equipment for department of ....., Nil Ratan Sircar Medical College, Kolkata-700014.

Ref:-

Sir,

Having examined the pre-qualification & other documents published in the N.I.T, I /we hereby submit all the necessary information and relevant documents for evaluation:

1. That the application is made by me / us on behalf of.....

In the capacity of.....duly authorized to submit the offer. The authorization letter from the Company is attached in Annexure II.

2. We accept the terms and conditions as laid down in the NIT mentioned above and declare that we shall abide by it for throughout the tender period and its extensions, if any.

3. We are offering rate for the following item /items and assured supply to the Health & Family Welfare Department, Government of West Bengal as per Table-1 (Instrument Specification)

4. a. We propose that the order and bill should be raised in our name.

For liasing, we have appointed M/S ..... having its office at ..... (Address, contact no and e mail address) as C&F agent /Liasoner /Contact person/..... Others (strike out whichever is not applicable) as per clause ..... of the NIT

b. The agreement between ourselves and the distributor/C&F/Liasoner& other documents as prescribed is attached in annexure III

5. In the event of being selected, supply will be made within the stipulated period excepting the condition which is beyond our control.

6. We understand that:

(a) Tender Selection Committee/ H&FW Dept can amend the scope & value of the contract bid under this project.

(b) Tender Selection Committee/ H&FW Dept reserves the right to reject any application without assigning any reason.

Date :-

Signature of applicant including title  
and capacity in which application is made.

Contact details -

Mobile:

E mail ID:



**Annexure II**

Authorization letter in favour of the applicant (other than Managing Director/ Proprietor) –  
from the competent authority.

FORMAT

(To be furnished in the Company's official letter pad with full address and contact no, E mail  
Address etc)

(TO WHOM IT MAY CONCERN)

This is to certify that Mr./Mrs. ....(Name),  
an employee of this Organisation as ..... (Official  
Designation) is hereby authorized to submit tender online , Vide NIT  
No....., Dated..... on behalf of the Organisation.

Signature of the  
competent authority with Seal

.....

(Signature of the Authorized Person)

Signature of Mr./Mrs. ....

.....is hereby attested.



## **Annexure- III**

- a. Copy of agreement between the Manufacturer and the Direct Importer/Distributor in Annexure -III



**ANNEXURE-IV**  
Draft Affidavit Proforma

I, Sri/Smt. ....  
The Managing Director/Proprietor (etc.) of the Firm.  
..... (Name of the firm)

At (address).....

P.O. ....

P.S.....Dist.....

do hereby solemnly affirm and declare as follows:

- a. That I am not under conviction of any offence making myself liable to be disqualified to supply of Surgical Equipments/Instruments etc. to any Govt. or Govt. undertaking Organization /Institution in the State of West Bengal or other State or States.
- b. My Firm has not been blacklisted at present as a whole or for any item/ items as quoted in this tender by any Govt. or Govt. undertaking Organization /Institution in the State of West Bengal or other State or States in India.
- c. That no case is pending against me or against my Firm in any criminal court of law in connection with the supply of Surgical/hospital equipment/Instruments to the Govt. or Govt. undertaking Organization / Institution in the State of West Bengal or other State or States ( If any case is pending, pl state the details ).
- d. *That, I declare that the item or items quoted by me confirms the specification given in the NIT document or higher standard relating to the specification is provided.*
- e. The duly authenticated Technical Data Sheet is uploaded conforming to the prescribed standard.
- f. That, I also declare that if any information subsequently found incorrect or false will it automatically render the tender submitted by me cancelled and make me liable for penal/legal action as per law of the country.
- g. That I do further affirm that the statements made by me in this tender are true to the best of my knowledge and belief and all the documents attached are genuine & correct.

Deponent



### Specification of HPLC

The modular HPLC system must be a quaternary gradient type, with upgradable with its own GUI controller comprises a colour LCD monitor and a Intuitive operation based touch-panel system and should be usable as pre-treatment system.

The internal walls of the module must be made of material that prevents corrosion due to the humidity and the vaporization of solvents that prevail in the system with a flow path design

All modules should have built in leak sensor

The interval of chromatography data station(CDS)' s single run direction to the autosampler response should be less than 20 seconds

UV-VIS detector System (wavelength range: 190 – 900nm, wavelength accuracy less than 2 nm) must be upgradable with PDA detector, 3-dimentional Fluorescence detector , RI detector and MS detector for future use.

Flow cell volume must be  $\leq 20 \mu\text{l}$  with  $\leq 15 \text{ mm}$  pathlength. Must capable of (i) automated peak purity check, (ii) Library search and (iii) display 3-D chromatogram

Suitable software must be able to control fully the HPLC system, all the upgradable modules such as pump, auto sampler, column oven, UV detector, DAD, fluorescence detector and RI detector

**Warranty: Three years**

*Soma*  
*20/12/18*

**Dr. Soma Gupta**  
Prof & Head  
Department of Biochemistry  
NRS Medical College  
Kolkata - 14

*[Signature]*  
*20/12/18*

**Prof. (Dr.) Tuphan Kanti Dolai**  
M.D. (Med.), D.N.B. (Med.), DM. (Clinical Haematology)  
Professor of Haematology Department  
NRSMC&H, Kolkata-14



### Specification of ELISA Reader

Sample Size :96 well format microplate (V-shaped, flat and round including strip wells )with  
Transparent bottom.

Absorbance :0.0-3.0 Abs

Resolution : 0.001 Abs

Precision : +/- 1.5% from 0.2 Abs

Detector : Silicon Photodiode

Wavelength : 6 position filter wheel with 405,450,490 and 630 nm filters supplied  
as standard accessory

Measurement : Monochromatic and Bi-chromatic mode

Light Source : Long life Halogen Lamp

Keyboard :Rugged, Waterproof, Membrane keypad

Display :Alphanumeric L.C.D.

Printer :High resolution Full Graphic Thermal Printer

Interface :RS -232 Serial Interface

Power Supply : 220 V+ 10 % , 50 Hz with On-line 1 KVA U.P. S. system .

**The ELISA micro plate reader should be interfaced and provided with a Computer,  
UPS, and Printer.**

**Three years warranty of the instrument is to be provided**

*Soma*  
20/12/18  
**Dr. Soma Gupta**  
Prof & Head  
Department of Biochemistry  
NRS Medical College  
Kolkata - 14

*Tuphan Kanti Dolai*  
20/12/18  
**Prof. (Dr.) Tuphan Kanti Dolai**  
M.D. (Med.), D.N.B. (Med.), B.M. (Clinical Haematology)  
Professor of Haematology Department  
NRSMC&H, Kolkata-14



### Specification of ELISA Washer

Volume of Solution dispensed: 50-3000  $\mu$ l steps for washing 50-400  $\mu$ l steps for dispensing

Accuracy @ 300  $\mu$ l :  $\pm$  10% well to well

Precision (Repeatability) @ 300  $\mu$ l : 5 % CV or better

Processing time : 350  $\mu$ l X 3 Cycles X 8 Port: 180 seconds

Residual Volume : > 5  $\mu$ l per well

Max suck time : 250 seconds

Wash Heads : Interchangeable 8 or 12 channel (Optional)

Programmes in the memory : 7 - 12

Liquid Contact Materials : Glass, Polypropylene, Polyethylene, Stainless Steel, Tygon,

Teflon, Delrin and Nylon

Recommended Decontaminants : Decron, Vikron or 5% Sodium Hypochlorite solution mixed with

De-ionised water.

Power requirements : 230 volt AC  $\pm$ 10% , 50/60 Hz

Power Consumption : 30 watts max

Operating Temperature : 10-40° C

Mandatory Pre-requisites : a) **Three years warranty of the instrument is to be provided**

*Handwritten signature*  
*20/12/18*  
**Dr. Soma Gupta**  
Prof & Head  
Department of Biochemistry  
NRS Medical College  
Kolkata - 14  
**Dr. Soma Gupta**  
Prof & Head  
Department of Biochemistry  
NRS Medical College  
Kolkata - 14

*Handwritten signature*  
*20/12/18*  
**Prof. (Dr.) Tuphan Kanti Dolai**  
M.D. (Med.), D.N.B. (Med.), D.M. (Clinical Haematology)  
Professor of Haematology Department  
NRS Medical College, Kolkata-14



**Specification of Fluorescence microscope**

Upright, able to detect GFP, RFP and yellow fluorescence protein, DAPI stain at magnification 10X, 40X and 100X, appropriate display systems and systems for capture and analysis of images

**Warranty: Three years**

*Soma*  
20/12/18

**Dr. Soma Gupta**  
Prof & Head  
Department of Biochemistry  
NRS Medical College  
Kolkata - 14

*Tuphan Kanti Dolai*  
20/12/18  
29/12/18

Prof. (Dr.) Tuphan Kanti Dolai  
M.D. (Med.), D.N.B. (Med.), D.M. (Clinical Haematology)  
Professor of Haematology Department  
NRSMC&H, Kolkata-14



NEW

VIDEO ENDOSCOPE	
Parameter /Feature	THRESHOLD LIMIT/RANGE
HD GASTROVIDEOSCOPE	
FIELD OF VIEW	140° OR MORE
DEPTH OF FIELD IN MM	(2-100) OR MORE
TIP DEFLECTION UP/DOWN IN DEG	210/90 OR MORE
TIP DEFLECTION LEFT/RIGHT IN DEG	100/100 OR MORE
RIGID DISTAL DIAMETER IN MM	9.2 OR LESS
INSERTION TUBE DIAMETER	9.2 mm OR LESS
INSTRUMENT CHANNEL	2.8 mm OR MORE
WORKING LENGTH	1030 mm OR MORE
TOTAL LENGTH	1350 mm OR MORE
HD VIDEO COLONOVIDEOSCOPE	
FIELD OF VIEW	140° OR MORE
DEPTH OF FIELD	2-100 mm OR MORE
TIP DEFLECTION UP/DOWN IN DEG	180/180 OR MORE
TIP DEFLECTION LEFT/RIGHT IN DEG	160/160 OR MORE
RIGID DISTAL DIAMETER IN MM	13 OR LESS
INSERTION TUBE DIAMETER IN MM	13 OR LESS
INSTRUMENT CHANNEL IN MM	3.7 OR MORE
WORKING LENGTH IN MM	1680 OR MORE
TOTAL LENGTH IN MM	2000 OR MORE
DUODENOVideoscope	
DIRECTION OF VIEW: BACKWARD SIDE	5° OR MORE
OUTER DIAMETER IN MM	13.7 OR LESS
CHANNEL DIAMETER IN MM	4.2 OR MORE
INSERTION TUBE LENGTH IN MM	1240 OR MORE

The Credentials and specifications have been fixed purely as per the need and requirements of Hospital / Medical College

*[Signature]*  
Dr. Gopal Krishna Dhalli  
Professor & Head  
School of Digestive & Liver Diseases  
IPGME&R  
Kolkata

*[Signature]*  
19/07/18

*[Signature]*  
HOD. G. Surgery  
M.C.U. 19.7.18

Partha Sarathi Patra  
Associate Professor  
Dept. of Gastroenterology  
SDLD IPGME&R SSKM



FIELD OF VIEW	100° OR MORE
DEPTH OF FIELD IN MM	(5-60) OR MORE
ANGULATIONS UP/DOWN	120°/90° OR MORE
ANGULATIONS RIGHT/LEFT	110°/90° OR MORE
VISIBLE DISTANCE	10 mm OR MORE
LIGHT SOURCE CUM VIDEO PROCESSOR	
COMPATIBILITY WITH INPUT VOLTAGE	(220 - 240) OR MORE
HD MEDICAL GRADE MONITOR SIZE IN INCH	21 OR MORE
MUST HAVE SPECIFICATIONS	
Parameter /Feature	Yes /No
HIGH DEFINITION GASTROVIDEOSCOPE	
LIGHT SOURCE: 300W XENON WITH EMERGENCY BACKUP	
FULLY IMMERSIBLE IN DISINFECTANT SOLUTION	
BUILT IN HD TV COMPATIBLE CCD WITH CLOSE FOCUS OBSERVATION CAPACITY	
SUITABLE FOR NARROW BAND IMAGING	
HD VIDEO COLONOVIDEOSCOPE	
FULLY IMMERSIBLE IN DISINFECTANT SOLUTION	
BUILT IN HD TV COMPATIBLE CCD WITH CLOSE FOCUS OBSERVATION CAPACITY	
SUITABLE FOR NARROW BAND IMAGING	
DUODENOVideoscope	
SUITABLE FOR NARROW BAND IMAGING	
LIGHT SOURCE CUM VIDEO PROCESSOR	
INTEGRATED WITH LIGHT SOURCE	
BAND IMAGING CAPABILITIES	
Quality Standard and Safety Certification	

*Gopal Krishna Dhali*  
**Dr. Gopal Krishna Dhali**  
 Professor & Head  
 School of Digestive & Liver Diseases  
 IPGME&R  
 Kolkata

*19/07/18*

*Sukumar Das*  
 HOD, G. Surgery  
 MEU 19.7.18

*Partha Sarathi Patra*  
 Associate Professor  
 Dept. of Gastroenterology SDL  
 IPGME&R, SSKM Hospital



Submission done  
HOD, A. Surgery  
on 19.7.18

19/07/18

Partha Sarathi Patra  
Associate Professor  
Dept. of Gastroenterology  
S.D.L.D I PGIMER SRM

Dr. Gopal Krishna Dhalli  
Professor & Head  
School of Digestive Liver Diseases  
IPGME&H  
Kolkata

Scanned by CamScanner

For the item Sl. No. 4 :

- 1.The scope should have auxiliary water jet channel & suitable for NBI/BLI/OE i-scan. Rigid distal diameter should be in 9.9 mm or more and insertion tube diameter should be in 9.8 mm or more.
- 2.Video processor should have facility of 300 W xenon or equivalent with emergency back up. It should have the facility of optical enhancement (NBI/BLI/OE i-scan). It must have HDTV medical grade monitor 24" or more, suitable trolley , software , computer & printer.

For the item Sl. No. 5 :

- 1.Scope should have auxiliary water jet channel & suitable for NBI/BLI/OE i-scan.
2. Video processor should have facility of 300 W xenon or equivalent with emergency back up. It should have the facility of optical enhancement (NBI/BLI/OE i-scan).



NEW

### ANAESTHESIA WORK STATION

Parameter / Feature	Threshold Limit / Range
Battery back up in minutes	60 or more
Gas flow in liter	(0.15 - 10) or more
Vaporiser flow range in liter/minutes	(0.2 - 15) or more
Vaporiser chamber capacity in ml.	250
Ventilator TFT display size (diagonal) in inch.	8 or more
Ventilator Tidal Volume in ml.	(20 - 1600) or more
Ventilator frequency in breaths/minutes	(4 - 50) or more
Ventilator I:E ratio	(1:1 - 1:3) or more
Ventilator inspiratory pause of Tidal Volume in %	0-60
Ventilator PEEP in cm H2O	(0-30) or more
Ventilator pressure limit in cm H2O	(0-70) or more
Sodalime capacity in liters	(1-2) or more
Drawers non lockable	2 or more
Provision for vaporisers	2 or more

### MUST HAVE SPECIFICATION

Parameter / Feature	Yes / No
Provision for delivery of oxygen, nitrus oxide and medical air	
Circle absorber, ventilator and vaporiser	
Independent attachments for connecting central gas and pin indexed cylinders	
Non interchangeable pipe line hose	
Large size pressure gauge	
Integrated independent oxygen flow meter	
Quick availability of gases on the time of machine on/off	
Selector switch should be there for open and close circuit	
The common gas outlet mounted at front of the machine	
Writing table/work surface	
Top shelf	
N2O cut off facility if o2 supply fails	
O2 failure alarm both visual and audible	
Availability of O2 flush bypassing vaporiser	
Visual display of individual gas.	
Maintenance free vaporiser for the life of the product	
Pneumatically driven and electronically controlled Ventilator	
Control of ventilator parameter by touch screen and rotary dial	
Availability of VCV, PCV, SIMV-V, spontaneous/pressure support, manual modes	
Flow triggered assist mode	
Leak compensation	
User adjustable alarms for major parameters	
User adjustable apnea alarms	
Display of patient spirometry loop and wave form (volume and air way pressure/flow)	
Display of compliance during Leak test	
Separate flow sensor for inspiratory and expiratory breathing path	
Disposable type volume measurement sensor	
Latex free and autoclavable breathing system except O2 sensor	
Upside down and clearly visible Bellows	
Provision for low flow anaesthesia	
10 galvanic oxygen sensor or paramagnetic in the inspiratory limb	
<b>Electrical Safety and Quality Certificaton</b>	
Standard & safety certification (European CE & US FDA approved)	
THE CREDENTIALS AND SPECIFICATIONS HAVE BEEN FIXED PURELY AS PER THE NEED AND REQUIREMENT OF HOSPITAL/MEDICAL COLLEGE	

*Syauki Mubtama*  
25.9.18

*Hin Kumm Kumdy*  
25.9.18

*BhallaChang*  
25/9/2018