


Government of West Bengal  
Office of the Principal  
NRS Medical College, Kolkata

Memo No: NMC/Store/ 10067

Date: 8/1/19

**Tender Notice**  
**(NIT NO. 61/2018-19)**

Sealed Tender is invited by the Principal, N. R. S. Medical College, Kolkata – 14 for supply of 1) **Vertical gel with blotting arrangement** Tender form and details/specification can be obtained from the Departmental website [www.wbhealth.gov.in](http://www.wbhealth.gov.in) from 09/01/2019. Last date of submission of Tender is 18/01/2019 up to 1 P.M. Tender will be opened on same date at 2.30 P.M in the office of the Principal, N. R. S. Medical College. Details are available in the Department of Health & Family Welfare, Govt. of West Bengal official Website: [www.wbhealth.gov.in](http://www.wbhealth.gov.in) & College website ([www.nrsmc.edu.in](http://www.nrsmc.edu.in))




Principal  
NRS Medical College,  
Kolkata-14

Memo No: NMC/Store/ 10067/(y)

Dated: 8/1/19

1. The Medical Superintendent-cum-Vice Principal, NRSMCH, Kolkata
2. Accounts Officer, NRS Medical College, Kolkata
3. I. T. Cell, Swasthya Bhawan, Salt-lake, for publication in the website.
4. Notice Board.



Principal  
NRS Medical College  
Kolkata-14



## Terms and Condition

1. Only Manufacturers/direct Importers/Authorized Distributors are eligible for quoting.
2. The rate should be quoted in both figure and word strictly as per sample/mentioned specification along with provided accounting unit only. All rates quoted should be inclusive of all liveable charges (including door delivery charges to the concerned department, packing charges, installation charges), and all other foreseeable incidental charges are exclusive of all applicable taxes should be mentioned separately.
3. The Tendering authority reserves the right to withdraw any item from the Tender at any stage. In such Contingency the selection of such item, if already made in favour of any Tendered shall be treated as cancelled. The Tendering authority also reserves the right to accept / reject / split any.
4. The Tenderer shall be bound for total replacement of the supplied quantity with a fresh supply of equivalent quantity if the item(s) are found to be defective/not as per given specification.
5. Payment will be made only after the delivery of the articles and satisfactory certification by an authority fixed by the administration. No request for advance or part payment will be considered.
6. In case of rejection of any supply on quality grounds or failure or refusal to replace the contracted items within the specified time or according to specifications/quality, the College/Hospital administration will have the authority to recover the losses, if any, by making deductions from any pending claim of the supplier in respect of same tender or any past/future tender.
7. The Specification and Brochure is to be submitted along with the Tender.
8. The Tenderers must accompany all the papers/documents along with the Tender.
9. The Tenderers must submit their tender in two bid method. Main envelope should contain Technical bid in Envelop A and Financial bid in Envelop B. all the envelopes must be properly sealed before submission.
10. Price bid of those bidders will only be opened who will qualify in technical scrutiny by tender selection committee.
11. Earnest money amounting to Rs. 5,000/- (Five thousand) only must be deposited in the form of Demand Draft from any Nationalized Bank in favour of 'Principal, N. R. S. Medical College' payable at Kolkata by L1 bidder before collection order.
12. The performance bank guarantee will be mandatory for all suppliers and will not be waived in any case. The successful renderers shall be required to furnish the 'performance bank guarantee' from any Nationalized/scheduled Bank acceptable to Govt. Of West Bengal for one year @ 10% of quoted base rate per item for which the tenderer has been selected as supplier.
13. The tendering authority has the liberty to ask for **performance demonstration** of any item during any stage of the tender process, before placing order for supply. In case of dissatisfaction, the authority will have the right to disapprove the offer of L1 bidder and go for the next bidder. Orders for supply of the approved products will be placed with the successful tenderers after execution of agreements and such supply shall have to be made in such installments as may be fixed or spread over the period to be specified in the supply orders to be made in pursuance of the agreements. The equipment is to be transported in such packaging so as to avoid damage to primary package of manufacturer. The successful tenderer will have to supply within the specified time schedule that had been assured at the time of selection as supplier.



**14. Guarantee / Warranty Period:**

The machine/Instrument should be covered by free on-site comprehensive Warranty for one year after successful installation. The tenderers must quote for 5 (Five) years' comprehensive on-site AMC of entire system (including all spares & labour) after the date of completion of free warranty period.

**15. Performance Bank Guarantee:**

The performance bank guarantee will be mandatory for all suppliers and will not be waived in any case. The successful tenderers shall be required to furnish the 'Performance Bank Guarantee' from any Nationalized/Scheduled Bank acceptable to Govt of West Bengal for One year @ 10% of quoted base rate per item for which the tenderer has been selected as supplier. The performance bank guarantee will have to be renewed after completion of Free Warranty period, for additional 5 (Five) years under the paid CAMC period before transfer of the CAMC amount to bidder's account.

16. In case of more than one L 1 bidder, weightage will be given to document number 7 of the following list.

**List of essential documents to be submitted**

1. Application submitted in Annexure- I
2. Copy of GST Registration certificate showing Registration No. (not provisional)
3. Copy of Valid Trade license
4. Copy of PAN Card of the bidder
5. Copies of Income Tax return for the last three years
6. Copies of Profit & Loss A/c and Balance Sheet for the last three years .
7. Credential of the bidder in successful supply to Govt. hospitals



Principal  
NRS Medical College  
Kol-14



## ANNEXURE-I

### APPLICATION FORMAT

(To be furnished in the Company's official letter pad with full address and contact no etc)

To The Principal, Nil Ratan Sircar Medical College, Kolkata-700014.

Sub: Tender No.-----Dt.------(Name of the Item -1)-----2)-----  
-----3)-----4)-----

----) for purchase and installation of equipment for department of ....., Nil Ratan Sircar Medical College, Kolkata-700014.

Ref:-

Sir,

Having examined the pre-qualification & other documents published in the N.I.T, I /we hereby submit all the necessary information and relevant documents for evaluation:

1. That the application is made by me / us on behalf of.....

In the capacity of.....duly authorized to submit the offer. The

authorization letter from the Company is attached in Annexure II.

2. We accept the terms and conditions as laid down in the NIT mentioned above and declare that we shall abide by it for throughout the tender period and its extensions, if any.

3. We are offering rate for the following item /items and assured supply to the Health & Family Welfare Department, Government of West Bengal as per Table-1 (Instrument Specification)

4. a. We propose that the order and bill should be raised in our name.

For liaoing, we have appointed M/S ..... having its office at  
..... (Address, contact no and e mail address) as  
C&F agent /Liasoner /Contact person/..... Others (strike out whichever is not applicable) as  
per clause ..... of the NIT

OR

b. We propose that order and bill should be raised in favour of our authorized distributor.

For that purpose, we have appointed M/S ..... having its office at  
..... (address with contact no and e mail address) as authorized Distributor  
who will receive order and payment in his name on our behalf.

c. The agreement between ourselves and the distributor/C&F/Liasoner & other documents  
as prescribed is attached in annexure III

5. In the event of being selected, supply will be made within the stipulated period  
excepting the condition which is beyond our control.

6. We understand that:

(a) Tender Selection Committee/ H&FW Dept can amend the scope & value of the contract bid  
under this project.

(b) Tender Selection Committee/ H&FW Dept reserves the right to reject any application  
without assigning any reason.

Date :-

Signature of applicant including title  
and capacity in which application is made.

Contact details -

Mobile:

E mail ID:



## Specification of Vertical gel with blotting arrangements

### Mini gel Specification- for 8 X 7 cm gels

High throughput- Capable of running up to 4 mini gel (8 X 7 Cm) simultaneously.

Should be supplied with the capability of running two gels.

Flexible- Capable of running hand cast as well as precast gel.

Running and casting module should be different

Interchangeable module- Should be capable of using blotting module to do western blotting.

Leak proof, tape free and easy assembly.

Patented Flap wing for leak proof assembly.

Permanently bonded spacer plates for leak proof, without agarose sealing & taping casting of gels.

Casting frame with simple cam closure mechanism that gives precision alignment on any flat surface.

Side by side casting stands that allow access to both gels simultaneously.

Patented colored sample loading guides to prevent the skipping or repeated loading lanes.

Modular design can be used do western blotting by using the blotting module only.

Should able to run gels in 15-20 mins.

Should come with buffer dam.

Should come with 10% Stainfree Fast Acrylamide

### Blotting Module:

It should be compatible with small SDS page Unit.

It should be capable of doing the western blotting of the mini gels.

It can run 2 blots simultaneously

Color coded cassettes for proper orientation of gels and membranes

System is able to transfer two mini gels.

Ice free cooling unit for proper temperature control for western blotting

Modular design can be used do electrophoresis by using the electrode assembly and caster only.

### Power Supply Basic-


Programmable power supply should be capable to operate four electrophoresis units simultaneously for four identical runs with graphic LED display.

The output range should be 10-300 V, 0.4-400 mA, 1-75 W.


Constant voltage, current or Power with Automatic crossover

Memory storage: 9 programs , 9 steps, Timer Control : 99 hr, 59 min

Automatic Power up after Power failure, Safety features: No-load detection; sudden load change detection.

  
20/12/18

**Dr. Soma Gupta**  
Prof & Head  
Department of Biochemistry  
NRS Medical College  
Kolkata - 14

  
21/12/18

**Prof. (Dr.) Tuptan Kanti Dolai**  
M.D. (Med.), D.N.B. (Med.), F.M. (Clinical Haematology)  
Professor of Haematology Department  
NRSMC&H, Kolkata-14